

QUALITY ASSURANCE / QUALITY CONTROL DATA

J

Preparation Batch ID: BOD101922C
Method Batch ID: MBOD101922C

Analysis Method: SM 5210 B

Preparation Type: No Prep
Preparation Date: 10/19/2022 3:00:00 PM

Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	-	% REC High Limit	%RPD	% RPD Limit
QA/QC Type: MB Lab Sample ID: BOD101922CMB Client Sample ID: BOD101922CMB Date Analyzed: 10/19/2022 3:00:00 PM												
Biochemical Oxygen Demand	1.0	1.0	2.0	U	mg/L							
QA/QC Type: LCS Lab Sample ID: BOD101922CLCS Client Sample ID: BOD101922CLCS Date Analyzed: 10/19/2022 3:00:00 PM												
Biochemical Oxygen Demand	100	100	216		mg/L	198	109	85	-	115		
QA/QC Type: LCSD Lab Sample ID: BOD101922CLCSD Client Sample ID: BOD101922CLCSD Date Analyzed: 10/19/2022 3:00:00 PM												
Biochemical Oxygen Demand	100	100	208		mg/L	198	105	85	-	115	3.8	20
QA/QC Type: DUP Lab Sample ID: BOD101922CDUP Client Sample ID: 302284DUP Date Analyzed: 10/19/2022 3:00:00 PM												
Biochemical Oxygen Demand	15	15	160		mg/L						6.1	20

Comments: